

Human ADAM10 (A Disintegrin And Metalloprotease 10) CLIA Kit | ADAM10 clia kit

Human ADAM10 (A Disintegrin And Metalloprotease 10) CLIA Kit

CATALOG #
AAA171145

REACTIVITY
Human


PRODUCT SYNONYMS

ADAM10 (A Disintegrin And Metalloprotease 10); Human ADAM10 (A Disintegrin And Metalloprotease 10) CLIA Kit; AD10; CD156c; CDw156; HsT18717; MADM; kuz; ADAM metallopeptidase domain 10; ADAM10 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes natural and recombinant ADAM10. No significant cross-reactivity or interference between ADAM10 and analogues was observed.
Assay Type	Sandwich
Detection Range	15.625-1000pg/mL
Sensitivity	Min: 9.375pg/mL; Max: 1000pg/mL

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ADAM10 clia kit

Intended Uses: This CLIA kit applies to the in vitro quantitative determination of ADAM10 concentrations in serum, plasma and other biological fluids.

Principle of the Assay: This kit uses Sandwich-CLIA as the method. The micro CLIA plate provided in this kit has been pre-coated with an antibody specific to ADAM10. Standards or samples are added to the appropriate micro CLIA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for ADAM10 and Avidin-Horseradish Peroxidase (HRP) conjugate is added to each micro plate well successively and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain ADAM10, biotinylated detection antibody and Avidin-HRP conjugate will appear fluorescence. The Relative light unit (RLU) value is measured spectrophotometrically by the Chemiluminescence immunoassay analyzer. The RLU value is positively associated with the concentration of ADAM10. You can calculate the concentration of ADAM10 in the samples by comparing the RLU value of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

