

Human Integrin Associated Protein ELISA Kit | IAP elisa kit

Human Integrin Associated Protein ELISA Kit

CATALOG #
AAA17564

REACTIVITY
Human


PRODUCT SYNONYMS

Integrin Associated Protein; Human Integrin Associated Protein ELISA Kit; IAP/Protein MER6/Antigenic surface determinant protein OA3/Protein name:Leukocyte surface antigen CD47/antigen identified by monoclonal 1D8/CD47 antigen (Rh-related antigen; integrin-associated signal transducer)/CD47 glycoprotein/CD47 molecule/integrin associated protein/Integrin-associated protein/leukocyte surface antigen CD47/MER6; IAP elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of IAP. No significant cross-reactivity or interference between IAP and analogues was observed.
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.156-10ng/ml
Sensitivity	0.094ng/ml
Intra-assay Precision	CV<8%
Inter-assay Precision	CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IAP elisa kit	Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Capture antibody was pre-coated onto 96-well plates. And the biotin conjugated antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the target amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of target can be calculated.
---	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/iap-human-elisa-kits/integrinassociatedprotein/17564>