

Human Cyclophilin A ELISA Kit | CYPA elisa kit

Human Cyclophilin A ELISA Kit

CATALOG #
AAA17590

REACTIVITY
Human

PRODUCT SYNONYMS

Cyclophilin A; Human Cyclophilin A ELISA Kit; CYPA/CYPH/PPIA/Rotamase/Cyclophilin A/Cyclosporin A-binding protein/CYPAMGC12404/CYPH/EC 5.2.1.8/MGC117158/MGC23397/peptidyl-prolyl cis-trans isomerase A/peptidylprolyl isomerase A (cyclophilin A)/PPIase A/Rotamase A/T cell cyclophilin; CYPA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of CYP3A4. No significant cross-reactivity or interference between CYP3A4 and analogues was observed.
Sequence Length	164
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	1.25-80ng/ml
Sensitivity	0.75ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard operation of a CYPA ELISA Kit are listed below. This standard curve was generated at our lab for demonstration purpose only. Users shall obtain standard curve as per experiment by themselves. (N/A=not applicable)

X	ng/ml	0	1.25	2.5	5	10	20	40	80
Y	OD450	0.115	0.439	0.717	1.295	1.658	2.083	2.393	2.609

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CYPA elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Capture antibody was pre-coated onto 96-well plates. And the biotin conjugated antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the target amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of target can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	318264336
NCBI GeneID	100304996

NCBI Accession #	NP_001187167.1
NCBI GenBank Nucleotide #	NM_001200238.1
NCBI Official Full Name	cyclophilin A
NCBI Official Symbol	LOC100304996
NCBI Protein Information	cyclophilin A

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/cypa-human-elisa-kits/cyclophilin-a/17590>