

## Human Sphingosine ELISA Kit | SPH elisa kit

Human Sphingosine (SPH) ELISA Kit

CATALOG #  
**AAA180651**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

Sphingosine; Human Sphingosine (SPH) ELISA Kit; SPH elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                 |
|-------------------------|-------------------------------------------------|
| Reactivity              | Human                                           |
| Specificity             | No cross-reaction with other factors.           |
| Samples                 | Human serum, plasma or cell culture supernatant |
| Assay Type              | Quantitative Sandwich                           |
| Detection Range         | 20 ng/ml-0.312 ng/ml                            |
| Sensitivity             | Up to 0.06 ng/ml                                |
| Intra-assay Precision   | <= 8%                                           |
| Inter-assay Precision   | <= 12%                                          |
| Preparation and Storage | Store all reagents at 2-8 degree C.             |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for SPH elisa kit

**Principle of the Assay:** This experiment use double-sandwich elisa technique and the ELISA Kit provided is typical. The pre-coated antibody is human SPH monoclonal antibody and the detecting antibody is polyclonal antibody with biotin labeled. Samples and biotin labeling antibody are added into ELISA plate wells and washed out with PBS or TBS. Then Avidin-peroxidase conjugates are added to ELISA wells in order; Use TMB substrate for coloring after reactant thoroughly washed out by PBS or TBS. TMB turns into blue in peroxidase catalytic and finally turns into yellow under the action of acid. The color depth and the testing factors in samples are positively correlated.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)

Online datasheet: <https://stage.aaabiotech.com/sph-human-elisa-kits/sphingosine/180651>