

Bovine Vitamin B1 (VB1) ELISA Kit | VB1 elisa kit

Bovine Vitamin B1 (VB1) ELISA Kit

CATALOG #
AAA181412

REACTIVITY
Bovine


PRODUCT SYNONYMS

Vitamin B1 (VB1); Bovine Vitamin B1 (VB1) ELISA Kit; VB1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Specificity	This kit recognizes MN in samples. No significant cross-reactivity or interference between MN and analogues was observed.
Samples	Serum, Plasma And Other Biological Fluids
Assay Type	Competitive
Detection Range	0.16-10ng/mL
Sensitivity	0.1ng/mL.
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level MN were tested 20 times on one plate, respectively
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level MN were tested on 3 different plates, 20 replicates in each plate.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VB1 elisa kit

Principle of the Assay: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with MN. During the reaction, MN in the sample or standard competes with a fixed amount of MN on the solid phase supporter for sites on the Biotinylated Detection Ab specific to MN. Excess conjugate and unbound sample or standard are washed from the plate, and Avidin conjugated to Horseradish Peroxidase (HRP) are added to each microplate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color change is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The concentration of MN in the samples is then determined by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/bovine-ovine-elisa-kits/vitamin-b1-vb1/181412>