

Porcine Matrix Metalloproteinase 1 (MMP1) ELISA Kit | MMP1 elisa kit

Pig Matrix Metalloproteinase 1 (MMP1) ELISA Kit

CATALOG #
AAA182837

REACTIVITY
Porcine


PRODUCT SYNONYMS

Matrix Metalloproteinase 1 (MMP1); Pig Matrix Metalloproteinase 1 (MMP1) ELISA Kit; CLGN; CLG1; Collagenase; Interstitial Collagenase; Vertebrate Collagenase; Fibroblast Collagenase; MMP1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Specificity	This assay has high sensitivity and excellent specificity for detection of MMP1. No significant cross-reactivity or interference between MMP1 and analogues was observed.
Sequence Length	512
Assay Type	Quantitative Sandwich
Samples	Porcine serum, plasma and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.115ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level MMP1 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level MMP1 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MMP1 elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to MMP1. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to MMP1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain MMP1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of MMP1 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of MMP1 in porcine serum, plasma and other biological fluids.
--	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	386768590
NCBI GeneID	37949
NCBI Accession #	NP_001246499.1
NCBI GenBank Nucleotide #	NM_001259570.2
NCBI Official Full Name	matrix metalloproteinase 1, isoform L
NCBI Official Synonym Full Names	Matrix metalloproteinase 1
NCBI Official Symbol	Mmp1
NCBI Official Synonym Symbols	CG4859; dm1-MMP; Dm1-MMP; DmelCG4859; dmmp1; dMMP1; l(2)k04809; Mmp 1; MMP-1; mmp1; MMp1; MMP1
NCBI Protein Information	CG4859 gene product from transcript CG4859-R1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/porcine-elisa-kits/matrix-metalloproteinase-1-mmp1/182837>