

Dog Collagen Type I Alpha 2 (COL1a2) ELISA Kit | COL1a2 elisa kit

Dog Collagen Type I Alpha 2 (COL1a2) ELISA Kit

CATALOG #
AAA183125

REACTIVITY
Dog


PRODUCT SYNONYMS

Collagen Type I Alpha 2 (COL1a2); Dog Collagen Type I Alpha 2 (COL1a2) ELISA Kit; COL1-A2; OI4; Collagen Alpha-2(I)chain; osteogenesis imperfecta type IV; Collagen Of Skin; Tendon And bone; Alpha-2 Chain; COL1a2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Dog
Specificity	This assay has high sensitivity and excellent specificity for detection of Collagen Type I Alpha 2 (COL1a2). No significant cross-reactivity or interference between Collagen Type I Alpha 2 (COL1a2) and analogues was observed.
Sequence Length	23
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and Other Biological Fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.126ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Collagen Type I Alpha 2 (COL1a2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Collagen Type I Alpha 2 (COL1a2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for COL1a2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Collagen Type I Alpha 2 (COL1a2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Collagen Type I Alpha 2 (COL1a2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Collagen Type I Alpha 2 (COL1a2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Collagen Type I Alpha 2 (COL1a2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	124265168
NCBI GeneID	1278
Molecular Weight	2,134 Da
NCBI Official Full Name	collagen type I alpha 2, partial
NCBI Official Synonym Full Names	collagen type I alpha 2 chain
NCBI Official Symbol	COL1A2
NCBI Official Synonym Symbols	OI4; EDSCV; EDSARTH2
NCBI Protein Information	collagen alpha-2(I) chain
UniProt Protein Name	Collagen type I alpha 2
UniProt Gene Name	COL1A2

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.