

Human Indoleamine-2,3-Dioxygenase (IDO) ELISA Kit | IDO elisa kit

Human Indoleamine-2,3-Dioxygenase (IDO) ELISA Kit

CATALOG #
AAA184809

REACTIVITY
Human


PRODUCT SYNONYMS

Indoleamine-2; 3-Dioxygenase (IDO); Human Indoleamine-2; 3-Dioxygenase (IDO) ELISA Kit; CD107B; INDO; Indoleamine-Pyrrole 2; 3 Dioxygenase; IDO elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of IDO. No significant cross-reactivity or interference between IDO and analogues was observed.
Sequence Length	453
Samples	Human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	0.312-20ng/mL
Sensitivity	<0.127ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level IDO were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level IDO were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IDO elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of IDO in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids. Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to IDO. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to IDO. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain IDO, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of IDO in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	1176339
NCBI GeneID	853541
UniProt Accession #	P47125; D6VWP8
Molecular Weight	50,775 Da
NCBI Official Full Name	Indoleamine 2,3-dioxygenase
NCBI Official Symbol	BNA2
NCBI Protein Information	dioxygenase BNA2
UniProt Protein Name	Indoleamine 2,3-dioxygenase
UniProt Gene Name	BNA2
UniProt Synonym Gene Names	IDO

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/indoleamine-23-dioxygenase-ido/184809>