

Human Orosomucoid 2 (ORM2) ELISA Kit | ORM2 elisa kit

Human Orosomucoid 2 (ORM2) ELISA Kit

CATALOG #
AAA185053

REACTIVITY
Human


PRODUCT SYNONYMS

Orosomucoid 2 (ORM2); Human Orosomucoid 2 (ORM2) ELISA Kit; AGP-B; AGP2; Alpha-1-Acid Glycoprotein Type 2; ORM2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Orosomucoid 2 (ORM2). No significant cross-reactivity or interference between Orosomucoid 2 (ORM2) and analogues was observed.
Sequence Length	201
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma and Other Biological Fluids
Detection Range	0.78-50ng/mL
Sensitivity	< 0.27ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Orosomucoid 2 (ORM2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Orosomucoid 2 (ORM2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ORM2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Orosomuroid 2 (ORM2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Orosomuroid 2 (ORM2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Orosomuroid 2 (ORM2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Orosomuroid 2 (ORM2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	119607823
NCBI GeneID	5005
UniProt Accession #	P19652; Q16571; Q5T538; Q6IB74; B2R5L2
Molecular Weight	23,603 Da
NCBI Official Full Name	orosomuroid 2
NCBI Official Synonym Full Names	orosomuroid 2
NCBI Official Symbol	ORM2
NCBI Official Synonym Symbols	AGP2; AGP-B; AGP-B'
NCBI Protein Information	alpha-1-acid glycoprotein 2
UniProt Protein Name	Alpha-1-acid glycoprotein 2
UniProt Gene Name	ORM2
UniProt Synonym Gene Names	AGP2; AGP 2; OMD 2

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/orosomuroid-2-orm2/185053>