

Rat Angiotensin I Converting Enzyme 2 (ACE2) ELISA Kit | ACE2 elisa kit

Rat Angiotensin I Converting Enzyme 2 (ACE2) ELISA Kit

CATALOG #
AAA185208

REACTIVITY
Rat


PRODUCT SYNONYMS

Angiotensin I Converting Enzyme 2 (ACE2); Rat Angiotensin I Converting Enzyme 2 (ACE2) ELISA Kit; ACEH; ACEII; ACE-II; Peptidyl-Dipeptidase A; ACE-related carboxypeptidase; Angiotensin-converting enzyme homolog; Metalloprotease MPROT15; ACE2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Angiotensin I Converting Enzyme 2 (ACE2). No significant cross-reactivity or interference between Angiotensin I Converting Enzyme 2 (ACE2) and analogues was observed.
Sequence Length	805
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue Homogenates and Other Biological Fluids
Detection Range	0.781-50ng/mL
Sensitivity	< 0.31ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Angiotensin I Converting Enzyme 2 (ACE2) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Angiotensin I Converting Enzyme 2 (ACE2) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ACE2 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Angiotensin I Converting Enzyme 2 (ACE2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Angiotensin I Converting Enzyme 2 (ACE2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Angiotensin I Converting Enzyme 2 (ACE2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Angiotensin I Converting Enzyme 2 (ACE2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	299482810
Molecular Weight	93,318 Da
NCBI Official Full Name	angiotensin I converting enzyme 2
UniProt Protein Name	Angiotensin-converting enzyme
UniProt Gene Name	ACE2

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/angiotensin-i-converting-enzyme-2-ace2/185208>