

Human Matrix Metalloproteinase 14 (MMP14) ELISA Kit | MMP14 elisa kit

Human Matrix Metalloproteinase 14 (MMP14) ELISA Kit

CATALOG #
AAA185456

REACTIVITY
Human


PRODUCT SYNONYMS

Matrix Metalloproteinase 14 (MMP14); Human Matrix Metalloproteinase 14 (MMP14) ELISA Kit; MMP-X1; MT1-MMP; MTMMP1; Membrane Inserted; Membrane-type matrix metalloproteinase 1; Membrane-type-1 matrix metalloproteinase; MMP14 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Matrix Metalloproteinase 14 (MMP14). No significant cross-reactivity or interference between Matrix Metalloproteinase 14 (MMP14) and analogues was observed.
Sequence Length	582
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue Homogenates and Other Biological Fluids
Detection Range	1.56-100ng/mL
Sensitivity	< 0.57ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Matrix Metalloproteinase 14 (MMP14) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Matrix Metalloproteinase 14 (MMP14) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MMP14 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Matrix Metalloproteinase 14 (MMP14). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Matrix Metalloproteinase 14 (MMP14). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Matrix Metalloproteinase 14 (MMP14), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Matrix Metalloproteinase 14 (MMP14) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	4826834
NCBI GeneID	4323
NCBI Accession #	NP_004986.1
NCBI GenBank Nucleotide #	NM_004995.3
UniProt Accession #	P50281; Q6GSF3; Q92678; A8K5L0
Molecular Weight	65,894 Da
NCBI Official Full Name	matrix metalloproteinase-14 preproprotein
NCBI Official Synonym Full Names	matrix metalloproteinase 14
NCBI Official Symbol	MMP14
NCBI Official Synonym Symbols	MMP-14; MMP-X1; MT-MMP; MT1MMP; MTMMP1; WNCHRS; MT1-MMP; MT-MMP 1
NCBI Protein Information	matrix metalloproteinase-14
UniProt Protein Name	Matrix metalloproteinase-14
UniProt Gene Name	MMP14
UniProt Synonym Gene Names	MMP-14; MT-MMP 1; MTMMP1; MT1-MMP; MT1MMP

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.