

Human Asporin (ASPEN) ELISA Kit | ASPEN elisa kit

Human Asporin (ASPEN) ELISA Kit

CATALOG #
AAA185666

REACTIVITY
Human


PRODUCT SYNONYMS

Asporin (ASPEN); Human Asporin (ASPEN) ELISA Kit; PLAP1; SLRR1C; Periodontal ligament-associated protein 1; ASPEN elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of asporin. No significant cross-reactivity or interference between asporin and analogues was observed.
Sequence Length	380
Assay Type	Quantitative Sandwich
Samples	Human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.126ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level asporin were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level asporin were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ASPN elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to asporin. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to asporin. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain asporin, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of asporin in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of asporin in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.
--	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	13625797
NCBI GeneID	54829
UniProt Accession #	Q9BXN1; Q5TBF3; Q96K79; Q96LD0; Q9NXP3
Molecular Weight	43,417 Da
NCBI Official Full Name	asporin
NCBI Official Synonym Full Names	asporin
NCBI Official Symbol	ASPN
NCBI Official Synonym Symbols	OS3; PLAP1; PLAP-1; SLRR1C
NCBI Protein Information	asporin
UniProt Protein Name	Asporin
UniProt Gene Name	ASPN
UniProt Synonym Gene Names	PLAP1; SLRR1C; PLAP-1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/asporin-aspn/185666>