

Human Serotonin Transporter (SERT) ELISA Kit | SERT elisa kit

Human Serotonin Transporter (SERT) ELISA Kit

CATALOG #
AAA185975

REACTIVITY
Human


PRODUCT SYNONYMS

Serotonin Transporter (SERT); Human Serotonin Transporter (SERT) ELISA Kit; SLC6A4; SLC6-A4; HTT; 5-HTT; 5HTT; OCD1; hSERT; Solute Carrier Family 6; Member 4; Neurotransmitter Transporter; Serotonin; Sodium-dependent serotonin transporter; SERT elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Serotonin Transporter (SERT). No significant cross-reactivity or interference between Serotonin Transporter (SERT) and analogues was observed.
Sequence Length	672
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates, Cell Lysates, Cell Culture Supernates and Other Biological Fluids
Detection Range	0.156-10ng/mL
Sensitivity	< 0.058ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Serotonin Transporter (SERT) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Serotonin Transporter (SERT) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SERT elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Serotonin Transporter (SERT). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Serotonin Transporter (SERT). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Serotonin Transporter (SERT), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Serotonin Transporter (SERT) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	58700442
NCBI GeneID	6532
UniProt Accession #	P31645; Q5EE02
Molecular Weight	74,979 Da
NCBI Official Full Name	serotonin transporter
NCBI Official Synonym Full Names	solute carrier family 6 member 4
NCBI Official Symbol	SLC6A4
NCBI Official Synonym Symbols	HTT; 5HTT; OCD1; SERT; 5-HTT; SERT1; hSERT; 5-HTTLPR
NCBI Protein Information	sodium-dependent serotonin transporter
UniProt Protein Name	Sodium-dependent serotonin transporter
UniProt Gene Name	SLC6A4
UniProt Synonym Gene Names	HTT; SERT; 5HTT

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.