

Human Cyclin Dependent Kinase 8 (CDK8) ELISA Kit | CDK8 elisa kit

Human Cyclin Dependent Kinase 8 (CDK8) ELISA Kit

CATALOG #
AAA186191

REACTIVITY
Human


PRODUCT SYNONYMS

Cyclin Dependent Kinase 8 (CDK8); Human Cyclin Dependent Kinase 8 (CDK8) ELISA Kit; K35; Cell division protein kinase 8; Mediator complex subunit CDK8; Mediator of RNA polymerase II transcription subunit CDK8; CDK8 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Cyclin Dependent Kinase 8 (CDK8). No significant cross-reactivity or interference between Cyclin Dependent Kinase 8 (CDK8) and analogues was observed.
Sequence Length	463
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates and Other Biological Fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.127ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Cyclin Dependent Kinase 8 (CDK8) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Cyclin Dependent Kinase 8 (CDK8) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CDK8 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cyclin Dependent Kinase 8 (CDK8). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Cyclin Dependent Kinase 8 (CDK8). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Cyclin Dependent Kinase 8 (CDK8), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Cyclin Dependent Kinase 8 (CDK8) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	4502745
NCBI GeneID	1024
NCBI Accession #	NP_001251.1
NCBI GenBank Nucleotide #	NM_001260.2
UniProt Accession #	P49336; Q5VUF3; Q6ISB5
Molecular Weight	53,156 Da
NCBI Official Full Name	cyclin-dependent kinase 8 isoform 1
NCBI Official Synonym Full Names	cyclin dependent kinase 8
NCBI Official Symbol	CDK8
NCBI Official Synonym Symbols	K35
NCBI Protein Information	cyclin-dependent kinase 8
UniProt Protein Name	Cyclin-dependent kinase 8
UniProt Gene Name	CDK8

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.