

Human Histone Deacetylase 6 (HDAC6) ELISA Kit | HDAC6 elisa kit

Human Histone Deacetylase 6 (HDAC6) ELISA Kit

CATALOG #
AAA186715

REACTIVITY
Human


PRODUCT SYNONYMS

Histone Deacetylase 6 (HDAC6); Human Histone Deacetylase 6 (HDAC6) ELISA Kit; HD6; JM21; HDAC6 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Histone Deacetylase 6 (HDAC6). No significant cross-reactivity or interference between Histone Deacetylase 6 (HDAC6) and analogues was observed.
Sequence Length	1,063
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates, Cell Lysates and Other Biological Fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.112ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Histone Deacetylase 6 (HDAC6) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Histone Deacetylase 6 (HDAC6) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HDAC6 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Histone Deacetylase 6 (HDAC6). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Histone Deacetylase 6 (HDAC6). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Histone Deacetylase 6 (HDAC6), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Histone Deacetylase 6 (HDAC6) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	13128864
NCBI GeneID	10013
NCBI Accession #	NP_006035.2
NCBI GenBank Nucleotide #	NM_006044.3
UniProt Accession #	Q9UBN7; O94975; Q6NT75; Q7L3E5; Q96CY0
Molecular Weight	114,361 Da
NCBI Official Full Name	histone deacetylase 6 isoform b
NCBI Official Synonym Full Names	histone deacetylase 6
NCBI Official Symbol	HDAC6
NCBI Official Synonym Symbols	HD6; JM21; CPBHM; PPP1R90
NCBI Protein Information	histone deacetylase 6
UniProt Protein Name	Histone deacetylase 6
UniProt Gene Name	HDAC6
UniProt Synonym Gene Names	KIAA0901; HD6

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.