

Human Claudin 8 (CLDN8) ELISA Kit | CLDN8 elisa kit

Human Claudin 8 (CLDN8) ELISA Kit

CATALOG #
AAA186788

REACTIVITY
Human


PRODUCT SYNONYMS

Claudin 8 (CLDN8); Human Claudin 8 (CLDN8) ELISA Kit; CLDN8 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Claudin 8 (CLDN8). No significant cross-reactivity or interference between Claudin 8 (CLDN8) and analogues was observed.
Sequence Length	225
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates, Cell Lysates and Other Biological Fluids
Detection Range	0.156-10ng/mL
Sensitivity	< 0.055ng/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Claudin 8 (CLDN8) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Claudin 8 (CLDN8) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CLDN8 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Claudin 8 (CLDN8). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Claudin 8 (CLDN8). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Claudin 8 (CLDN8), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Claudin 8 (CLDN8) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	40788011
NCBI GeneID	9073
NCBI Accession #	NP_955360.1
NCBI GenBank Nucleotide #	NM_199328.2
UniProt Accession #	P56748; Q53EX7; D3DSE3
Molecular Weight	24,845 Da
NCBI Official Full Name	claudin-8
NCBI Official Synonym Full Names	claudin 8
NCBI Official Symbol	CLDN8
NCBI Official Synonym Symbols	HEL-S-79
NCBI Protein Information	claudin-8
UniProt Protein Name	Claudin-8
UniProt Gene Name	CLDN8

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/claudin-8-cldn8/186788>