

Human Hyaluronan Synthase 2 (HAS2) ELISA Kit | HAS2 elisa kit

Human Hyaluronan Synthase 2 (HAS2) ELISA Kit

CATALOG #
AAA186848

REACTIVITY
Human


PRODUCT SYNONYMS

Hyaluronan Synthase 2 (HAS2); Human Hyaluronan Synthase 2 (HAS2) ELISA Kit; Hyaluronate synthase 2; Hyaluronic acid synthase 2; HAS2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of HAS2. No significant cross-reactivity or interference between HAS2 and analogues was observed.
Sequence Length	552
Assay Type	Quantitative Sandwich
Samples	Human serum, plasma, tissue homogenates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	< 0.057ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level HAS2 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level HAS2 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for HAS2 elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to HAS2. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to HAS2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain HAS2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of HAS2 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of HAS2 in human serum, plasma, tissue homogenates and other biological fluids.
--	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	4885391
NCBI GeneID	3037
NCBI Accession #	NP_005319.1
NCBI GenBank Nucleotide #	NM_005328.2
UniProt Accession #	Q92819; Q32MM3
Molecular Weight	63,566 Da
NCBI Official Full Name	hyaluronan synthase 2
NCBI Official Synonym Full Names	hyaluronan synthase 2
NCBI Official Symbol	HAS2
NCBI Protein Information	hyaluronan synthase 2
UniProt Protein Name	Hyaluronan synthase 2
UniProt Gene Name	HAS2
UniProt Synonym Gene Names	HA synthase 2

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/hyaluronan-synthase-2-has2/186848>