

Human Semaphorin 3C (SEMA3C) ELISA Kit | SEMA3C elisa kit

Human Semaphorin 3C (SEMA3C) ELISA Kit

CATALOG #
AAA187774

REACTIVITY
Human


PRODUCT SYNONYMS

Semaphorin 3C (SEMA3C); Human Semaphorin 3C (SEMA3C) ELISA Kit; SEMAE; SemE; Semaphorin-E; Sema Domain; Immunoglobulin Domain(Ig); Transmembrane Domain(TM)and Short Cytoplasmic Domain 3C; SEMA3C elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Semaphorin 3C (SEMA3C). No significant cross-reactivity or interference between Semaphorin 3C (SEMA3C) and analogues was observed.
Sequence Length	233
Assay Type	Double-antibody Sandwich
Samples	Serum, Plasma, Tissue Homogenates and Other Biological Fluids
Detection Range	15.6-1000pg/mL
Sensitivity	< 6.7pg/mL
Application	Enzyme-linked immunosorbent assay for Antigen Detection.
Intra-assay Precision (Precision within an assay)	3 samples with low, middle and high level Semaphorin 3C (SEMA3C) were tested 20 times on one plate, respectively.
Inter-assay Precision (Precision between assays)	3 samples with low, middle and high level Semaphorin 3C (SEMA3C) were tested on 3 different plates, 8 replicates in each plate.
CV(%) =	SD/meanX100
Intra-Assay	CV<10%
Inter-Assay	CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SEMA3C elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Semaphorin 3C (SEMA3C). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Semaphorin 3C (SEMA3C). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Semaphorin 3C (SEMA3C), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Semaphorin 3C (SEMA3C) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	5454048
NCBI GeneID	10512
NCBI Accession #	NP_006370.1
NCBI GenBank Nucleotide #	NM_006379.4
UniProt Accession #	Q99985; B4DRL8
Molecular Weight	26,255 Da
NCBI Official Full Name	semaphorin-3C isoform 2
NCBI Official Synonym Full Names	semaphorin 3C
NCBI Official Symbol	SEMA3C
NCBI Official Synonym Symbols	SemE; SEMAE
NCBI Protein Information	semaphorin-3C
UniProt Protein Name	Semaphorin-3C
UniProt Gene Name	SEMA3C
UniProt Synonym Gene Names	SEMAE; Sema E

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.