

Human Butyrophilin Subfamily 3, Member A1 (BTN3A1) ELISA Kit | BTN3A1 elisa kit

Human Butyrophilin Subfamily 3, Member A1 (BTN3A1) ELISA Kit

CATALOG #
AAA188299

REACTIVITY
Human


PRODUCT SYNONYMS

Butyrophilin Subfamily 3; Member A1 (BTN3A1); Human Butyrophilin Subfamily 3; Member A1 (BTN3A1) ELISA Kit; CD277; BT3.1; BTF5; BTN3A1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Butyrophilin Subfamily 3, Member A1 (BTN3A1). No significant cross-reactivity or interference between Butyrophilin Subfamily 3, Member A1 (BTN3A1) and analogues was observed.
Sequence Length	513
Assay Type	Double-antibody Sandwich
Samples	Tissue Homogenates, Cell Lysates and Other Biological Fluids
Detection Range	0.312-20ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.117ng/mL
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Butyrophilin Subfamily 3, Member A1 (BTN3A1) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Butyrophilin Subfamily 3, Member A1 (BTN3A1) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for BTN3A1 elisa kit	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Butyrophilin Subfamily 3, Member A1 (BTN3A1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Butyrophilin Subfamily 3, Member A1 (BTN3A1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Butyrophilin Subfamily 3, Member A1 (BTN3A1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Butyrophilin Subfamily 3, Member A1 (BTN3A1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	37595558
NCBI GeneID	11119
NCBI GenBank Nucleotide #	NP_008979.3
NCBI Official Full Name	butyrophilin subfamily 3 member A1 isoform a
NCBI Official Synonym Full Names	butyrophilin subfamily 3 member A1
NCBI Official Symbol	BTN3A1
NCBI Official Synonym Symbols	BT5F; BT3.1; CD277; BTN3.1
NCBI Protein Information	butyrophilin subfamily 3 member A1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/butyrophilin-subfamily-3-member-a1-btn3a1/188299>