

Human Complement 1q (C1q) ELISA Kit | C1q elisa kit

Human Anti-Complement 1q Antibody (Anti-C1q) ELISA Kit


CATALOG # AAA188362	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------

PRODUCT SYNONYMS

Complement 1q (C1q); Human Anti-Complement 1q Antibody (Anti-C1q) ELISA Kit; Complement 1q; C1-q; Autoantigen Clq; C1q elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Anti-Complement 1q Antibody (Anti-C1q). No significant cross-reactivity or interference between Anti-Complement 1q Antibody (Anti-C1q) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	1.56-100ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.61ng/mL
Applicable Applications for C1q elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Anti-Complement 1q Antibody (Anti-C1q) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Anti-Complement 1q Antibody (Anti-C1q) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Test Principle	The microtiter plate provided in this kit has been pre-coated with an antigen. Standards or samples are then added to the appropriate microtiter plate wells with a Horseradish Peroxidase (HRP)-conjugated secondary antibody. After TMB substrate solution is added, those wells that contain Anti-Complement 1q Antibody (Anti-C1q) will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of Anti-Complement 1q Antibody (Anti-C1q) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/complement-1q-c1q/188362>