

Human Mannose Binding Lectin (MBL) ELISA Kit | MBL elisa kit

Human Anti-Mannose Binding Lectin Antibody (Anti-MBL) ELISA Kit

CATALOG # AAA188371	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------


PRODUCT SYNONYMS

Mannose Binding Lectin (MBL); Human Anti-Mannose Binding Lectin Antibody (Anti-MBL) ELISA Kit; Mannose Binding Lectin; MBL2; COLEC1; HSMBPC; MBP1; MBP; Collectin-1; Mannose-binding protein C; Mannan Binding Protein; Mannose-Binding Lectin(Protein C)2; Soluble(Opsonic Defect); MBL elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Anti-Mannose Binding Lectin Antibody (Anti-MBL). No significant cross-reactivity or interference between Anti-Mannose Binding Lectin Antibody (Anti-MBL) and analogues was observed.
Sequence Length	248
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	6.25-400ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 2.45ng/mL
Applicable Applications for MBL elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Anti-Mannose Binding Lectin Antibody (Anti-MBL) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Anti-Mannose Binding Lectin Antibody (Anti-MBL) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microtiter plate provided in this kit has been pre-coated with an antigen. Standards or samples are then added to the appropriate microtiter plate wells with a Horseradish Peroxidase (HRP)-conjugated secondary antibody. After TMB substrate solution is added, those wells that contain Anti-Mannose Binding Lectin Antibody (Anti-MBL) will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of Anti-Mannose Binding Lectin Antibody (Anti-MBL) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	4557739
NCBI GeneID	4153
NCBI Accession #	NP_000233.1
NCBI GenBank Nucleotide #	NP_000233.1
UniProt Accession #	P11226
NCBI Official Full Name	mannose-binding protein C
NCBI Official Synonym Full Names	mannose binding lectin 2
NCBI Official Symbol	MBL2
NCBI Official Synonym Symbols	MBL; MBP; MBP1; MBPD; MBL2D; MBP-C; COLEC1; HSMBPC
NCBI Protein Information	mannose-binding protein C
UniProt Protein Name	Mannose-binding protein C
UniProt Gene Name	MBL2
UniProt Synonym Gene Names	COLEC1; MBL; MBP-C
UniProt Entry Name	MBL2_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.