

Canine Immunoglobulin G (IgG) CLIA Kit | IgG clia kit

Canine Immunoglobulin G (IgG) CLIA Kit

CATALOG # AAA188412	REACTIVITY Canine	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	-----------------------------	---


PRODUCT SYNONYMS

Immunoglobulin G (IgG); Canine Immunoglobulin G (IgG) CLIA Kit; Immunoglobulin G; IgG clia kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Specificity	This assay has high sensitivity and excellent specificity for detection of Immunoglobulin G (IgG). No significant cross-reactivity or interference between Immunoglobulin G (IgG) and analogues was observed.
Sequence Length	205
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	1.17-300ug/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.52ug/mL
Applicable Applications for IgG clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Immunoglobulin G (IgG) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Immunoglobulin G (IgG) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to Immunoglobulin G (IgG). A competitive inhibition reaction is launched between biotin labeled Immunoglobulin G (IgG) and unlabeled Immunoglobulin G (IgG) (Standards or samples) with the pre-coated antibody specific to Immunoglobulin G (IgG). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Immunoglobulin G (IgG) in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the Immunoglobulin G (IgG) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	1389716324
NCBI Official Full Name	immunoglobulin G, partial

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/canine-clia-kits/immunoglobulin-g-igg/188412>