

Human Apolipoprotein M (APOM) ELISA Kit | APOM elisa kit

Human Apolipoprotein M (APOM) High Sensitive ELISA Kit

CATALOG # AAA188580	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------


PRODUCT SYNONYMS

Apolipoprotein M (APOM); Human Apolipoprotein M (APOM) High Sensitive ELISA Kit; Apolipoprotein M; Apo-M; G3a; NG20; Protein G3a; APOM elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of High Sensitive Apolipoprotein M (APOM). No significant cross-reactivity or interference between High Sensitive Apolipoprotein M (APOM) and analogues was observed.
Sequence Length	116
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.121ng/mL
Applicable Applications for APOM elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level High Sensitive Apolipoprotein M (APOM) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level High Sensitive Apolipoprotein M (APOM) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to High Sensitive Apolipoprotein M (APOM). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to High Sensitive Apolipoprotein M (APOM). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain High Sensitive Apolipoprotein M (APOM), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of High Sensitive Apolipoprotein M (APOM) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	371873523
NCBI GeneID	55937
NCBI Accession #	NP_001243098.1
NCBI GenBank Nucleotide #	NP_001243098.1
UniProt Accession #	O95445
NCBI Official Full Name	apolipoprotein M isoform 2
NCBI Official Synonym Full Names	apolipoprotein M
NCBI Official Symbol	APOM
NCBI Official Synonym Symbols	G3a; NG20; apo-M; HSPC336
NCBI Protein Information	apolipoprotein M
UniProt Protein Name	Apolipoprotein M
UniProt Gene Name	APOM
UniProt Synonym Gene Names	G3A; NG20; Apo-M; ApoM
UniProt Entry Name	APOM_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.