

Rat Coagulation Factor XII (F12) ELISA Kit | F12 elisa kit

Rat Coagulation Factor XII (F12) Mini Samples ELISA Kit

CATALOG # AAA188712	REACTIVITY Rat	APPLICATIONS ELISA
-------------------------------	--------------------------	------------------------------


PRODUCT SYNONYMS

Coagulation Factor XII (F12); Rat Coagulation Factor XII (F12) Mini Samples ELISA Kit; Coagulation Factor XII; FXII; HAF; Contact Factor; Hageman Factor; F12 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Mini Samples Coagulation Factor XII (F12). No significant cross-reactivity or interference between Mini Samples Coagulation Factor XII (F12) and analogues was observed.
Sequence Length	237
Assay Type	Double-antibody Sandwich
Samples	Plasma
Detection Range	0.625-40ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.232ng/mL
Applicable Applications for F12 elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Mini Samples Coagulation Factor XII (F12) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Mini Samples Coagulation Factor XII (F12) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Mini Samples Coagulation Factor XII (F12). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Mini Samples Coagulation Factor XII (F12). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Mini Samples Coagulation Factor XII (F12), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of Mini Samples Coagulation Factor XII (F12) in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	903736
NCBI GeneID	2161
NCBI Official Full Name	coagulation factor XII, partial
NCBI Official Synonym Full Names	coagulation factor XII
NCBI Official Symbol	F12
NCBI Official Synonym Symbols	HAF; HAE3; HAEX
NCBI Protein Information	coagulation factor XII

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/coagulation-factor-xii-f12/188712>