

Rat Fibrinopeptide A (FPA) ELISA Kit | FPA elisa kit

Rat Fibrinopeptide A (FPA) Mini Samples ELISA Kit


CATALOG # AAA188734	REACTIVITY Rat	APPLICATIONS ELISA
-------------------------------	--------------------------	------------------------------

PRODUCT SYNONYMS

Fibrinopeptide A (FPA); Rat Fibrinopeptide A (FPA) Mini Samples ELISA Kit; Fibrinopeptide A; FP-A; FPA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Mini Samples Fibrinopeptide A (FPA). No significant cross-reactivity or interference between Mini Samples Fibrinopeptide A (FPA) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	370.4-30,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 129.1pg/mL
Applicable Applications for FPA elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Mini Samples Fibrinopeptide A (FPA) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Mini Samples Fibrinopeptide A (FPA) were tested on 3 different plates, 8 replicates in each plate. $CV(\%) = SD/mean \times 100$ Intra-Assay: $CV < 10\%$ Inter-Assay: $CV < 12\%$
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Mini Samples Fibrinopeptide A (FPA) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Mini Samples Fibrinopeptide A (FPA) and unlabeled Mini Samples Fibrinopeptide A (FPA) (Standards or samples) with the pre-coated antibody specific to Mini Samples Fibrinopeptide A (FPA). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Mini Samples Fibrinopeptide A (FPA) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Mini Samples Fibrinopeptide A (FPA) in the sample.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/fibrinopeptide-a-fpa/188734>