

Rat Myeloperoxidase (MPO) CLIA Kit | MPO clia kit

Rat Myeloperoxidase (MPO) CLIA Kit

CATALOG # AAA188958	REACTIVITY Rat	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	--------------------------	---


PRODUCT SYNONYMS

Myeloperoxidase (MPO); Rat Myeloperoxidase (MPO) CLIA Kit; Myeloperoxidase; MPO clia kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Myeloperoxidase (MPO). No significant cross-reactivity or interference between Myeloperoxidase (MPO) and analogues was observed.
Sequence Length	458
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates and other biological fluids
Detection Range	68.6-50,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 29.8pg/mL
Applicable Applications for MPO clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Myeloperoxidase (MPO) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Myeloperoxidase (MPO) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Myeloperoxidase (MPO). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Myeloperoxidase (MPO). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Myeloperoxidase (MPO) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	157820285
NCBI GeneID	303413
NCBI Accession #	NP_001100506.1
NCBI GenBank Nucleotide #	NP_001100506.1
UniProt Accession #	D3ZGE2
NCBI Official Full Name	myeloperoxidase
NCBI Official Synonym Full Names	myeloperoxidase
NCBI Official Symbol	Mpo
NCBI Protein Information	myeloperoxidase
UniProt Protein Name	Myeloperoxidase
UniProt Gene Name	Mpo

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-clia-kits/myeloperoxidase-mpo/188958>