

Rat Prothrombin Fragment 1+2 (F1+2) CLIA Kit | F2 clia kit

Rat Prothrombin Fragment 1+2 (F1+2) CLIA Kit

CATALOG # AAA188971	REACTIVITY Rat	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	--------------------------	---


PRODUCT SYNONYMS

Prothrombin Fragment 1+2 (F1+2); Rat Prothrombin Fragment 1+2 (F1+2) CLIA Kit; Prothrombin Fragment 1+2; F2 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Prothrombin Fragment 1+2 (F1+2). No significant cross-reactivity or interference between Prothrombin Fragment 1+2 (F1+2) and analogues was observed.
Sequence Length	617
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.
Detection Range	0.55-400ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.23ng/mL
Applicable Applications for F2 clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Prothrombin Fragment 1+2 (F1+2) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Prothrombin Fragment 1+2 (F1+2) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	Average the duplicate readings for each standard, control, and samples and subtract the average zero standard relative light unit (RLU). Create a standard curve on log-log graph paper, with F1+2 concentration on the y-axis and the RLU value on the x-axis. Draw the best fit straight line through the standard points and it can be determined by regression analysis. Using some plot software, such as curve expert 1.30, is also recommended. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	135809
NCBI GeneID	29251
UniProt Accession #	P18292
NCBI Official Full Name	Prothrombin
NCBI Official Synonym Full Names	coagulation factor II
NCBI Official Symbol	F2
NCBI Protein Information	prothrombin
UniProt Protein Name	Prothrombin
UniProt Gene Name	F2
UniProt Entry Name	THRB_RAT

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-clia-kits/prothrombin-fragment-12-f12/188971>