

Human Programmed Cell Death Protein 1 Ligand 1 (PDL1) CLIA Kit | PDL1 clia kit

Human Programmed Cell Death Protein 1 Ligand 1 (PDL1) CLIA Kit

CATALOG # AAA188983	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---


PRODUCT SYNONYMS

Programmed Cell Death Protein 1 Ligand 1 (PDL1); Human Programmed Cell Death Protein 1 Ligand 1 (PDL1) CLIA Kit; Programmed Cell Death Protein 1 Ligand 1; CD274; PDCD1LG1; B7-H; B7H1; PD-L1; PDCD1L1; PDCD1-LG1; Programed Death Ligand 1; PDL1 clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Programmed Cell Death Protein 1 Ligand 1 (PDL1). No significant cross-reactivity or interference between Programmed Cell Death Protein 1 Ligand 1 (PDL1) and analogues was observed.
Sequence Length	176
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates and other biological fluids
Detection Range	27.4-20,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 11.5pg/mL
Applicable Applications for PDL1 clia kit	CLIA (Chemiluminescence Immunoassay)
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Programmed Cell Death Protein 1 Ligand 1 (PDL1) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Programmed Cell Death Protein 1 Ligand 1 (PDL1) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	The microplate provided in this kit has been pre-coated with an antibody specific to Programmed Cell Death Protein 1 Ligand 1 (PDL1). Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to Programmed Cell Death Protein 1 Ligand 1 (PDL1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is proportional to the Programmed Cell Death Protein 1 Ligand 1 (PDL1) level in the sample or standard.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

NCBI AND UNIPROT INFORMATION

NCBI GI #	390979639
NCBI GeneID	29126
NCBI Accession #	NP_001254635.1
NCBI GenBank Nucleotide #	NP_001254635.1
UniProt Accession #	Q9NZQ7
NCBI Official Full Name	programmed cell death 1 ligand 1 isoform b
NCBI Official Synonym Full Names	CD274 molecule
NCBI Official Symbol	CD274
NCBI Official Synonym Symbols	B7-H; B7H1; PDL1; PD-L1; hPD-L1; PDCD1L1; PDCD1LG1
NCBI Protein Information	programmed cell death 1 ligand 1
UniProt Protein Name	Programmed cell death 1 ligand 1
UniProt Gene Name	CD274
UniProt Synonym Gene Names	B7H1; PDCD1L1; PDCD1LG1; PDL1; PD-L1; PDCD1 ligand 1; Programmed death ligand 1; B7-H1
UniProt Entry Name	PD1L1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/programmed-cell-death-protein-1-ligand-1-pdl1/188983>