

Rat D-Dimer (D2D) ELISA Kit | D2D elisa kit

Rat D-Dimer (D2D) Wide-range ELISA Kit


CATALOG # AAA189165	REACTIVITY Rat	APPLICATIONS ELISA
-------------------------------	--------------------------	------------------------------

PRODUCT SYNONYMS

D-Dimer (D2D); Rat D-Dimer (D2D) Wide-range ELISA Kit; D-Dimer; D 2 Dimer; D2D elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Wide-range D-Dimer (D2D). No significant cross-reactivity or interference between Wide-range D-Dimer (D2D) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Plasma
Detection Range	61.7-5,000ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 27.2ng/mL
Applicable Applications for D2D elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Wide-range D-Dimer (D2D) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Wide-range D-Dimer (D2D) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Wide-range D-Dimer (D2D) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Wide-range D-Dimer (D2D) and unlabeled Wide-range D-Dimer (D2D) (Standards or samples) with the pre-coated antibody specific to Wide-range D-Dimer (D2D). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Wide-range D-Dimer (D2D) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Wide-range D-Dimer (D2D) in the sample.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/d-dimer-d2d/189165>