

Human Salusin Beta (SALb) ELISA Kit | SALb elisa kit

Human Salusin Beta (SALb) Wide-range ELISA Kit


CATALOG # AAA189200	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------

PRODUCT SYNONYMS

Salusin Beta (SALb); Human Salusin Beta (SALb) Wide-range ELISA Kit; Salusin Beta; SALb elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Wide-range Salusin Beta (SALb). No significant cross-reactivity or interference between Wide-range Salusin Beta (SALb) and analogues was observed.
Assay Type	Competitive Inhibition
Samples	Serum, plasma and other biological fluids
Detection Range	24.69-2,000pg/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 8.74pg/mL
Applicable Applications for SALb elisa kit	ELISA
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Wide-range Salusin Beta (SALb) were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Wide-range Salusin Beta (SALb) were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Test Principle	This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to Wide-range Salusin Beta (SALb) has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled Wide-range Salusin Beta (SALb) and unlabeled Wide-range Salusin Beta (SALb) (Standards or samples) with the pre-coated antibody specific to Wide-range Salusin Beta (SALb). After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of Wide-range Salusin Beta (SALb) in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of Wide-range Salusin Beta (SALb) in the sample.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.



Standard Curve (Sample)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/salusin-beta-salb/189200>