

Human Ischemia Modified Albumin (IMA) CLIA Kit | IMA clia kit

Ischemia Modified Albumin (IMA) CLIA Kit

CATALOG # AAA189245	REACTIVITY Human	APPLICATIONS CLIA (Chemiluminescence Immunoassay)
-------------------------------	----------------------------	---


PRODUCT SYNONYMS

Ischemia Modified Albumin (IMA); Ischemia Modified Albumin (IMA) CLIA Kit; IMA clia kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of IMA. No significant cross-reactivity or interference between IMA and analogues was observed.
Samples	Serum, plasma and other biological fluids
Assay Type	Quantitative Competitive
Detection Range	370.4-30,000ng/mL
Sensitivity	< 131.3ng/mL
Applicable Applications for IMA clia kit	CLIA (Chemiluminescence Immunoassay)
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level IMA were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level IMA were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IMA clia kit	Intended Uses: The Chemiluminescent Immunoassay kit is designed for the in vitro sensitive quantitative measurement of IMA in human serum, plasma and other biological fluids. Principle of the Assay: The microplate provided in this kit has been pre-coated with a monoclonal antibody specific to IMA. A competitive inhibition reaction is launched between biotin labeled IMA and unlabeled IMA (Standards or samples) with the pre-coated antibody specific to IMA. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of IMA in the sample. Then the mixture of substrate A and B is added to generate glow light emission kinetics. Upon plate development, the intensity of the emitted light is reverse proportional to the IMA level in the sample or standard.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-clia-kits/ischemia-modified-albumin-ima/189245>