

Human Bcl2 Associated Death Promoter (BAD) ELISA Kit | BAD elisa kit

Bcl2 Associated Death Promoter (BAD) ELISA Kit

CATALOG # AAA189380	REACTIVITY Human	APPLICATIONS ELISA
-------------------------------	----------------------------	------------------------------


PRODUCT SYNONYMS

Bcl2 Associated Death Promoter (BAD); Bcl2 Associated Death Promoter (BAD) ELISA Kit; BBC2; BCL2L8; BCL2-Antagonist Of Cell Death; Bcl-2-binding component 6; Bcl-2-like protein 8; Bcl-xL/Bcl-2-associated death promoter; Bcl2 antagonist of cell death; BAD elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of BAD. No significant cross-reactivity or interference between BAD and analogues was observed.
Samples	Human tissue homogenates, cell lysates and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	78-5,000pg/mL
Sensitivity	< 31pg/mL
Applicable Applications for BAD elisa kit	ELISA
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level BAD were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level BAD were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for BAD elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of BAD in human tissue homogenates, cell lysates and other biological fluids. Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to BAD. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to BAD. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain BAD, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of BAD in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/bcl2-associated-death-promoter-bad/189380>