

Uric Acid (UA) ELISA Kit | UA elisa kit

General species Uric Acid (UA) ELISA Kit

CATALOG #
AAA189663

REACTIVITY
General species


PRODUCT SYNONYMS

Uric Acid (UA); General species Uric Acid (UA) ELISA Kit; Urate; UA/Uric Acid; UA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General species
Specificity	This assay has high sensitivity and excellent specificity for detection of UA. No significant cross-reactivity or interference between UA and analogues was observed.
Assay Type	Competitive
Samples	Serum, plasma and other biological fluids
Detection Range	617.3-50,000ng/mL
Sensitivity	< 243.1ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level UA were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level UA were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	Store at 4 degree C for 1 month/-20 degree C for long-term (within shelf life) Shelf Life: 8 months The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for UA elisa kit	Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to UA has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled UA and unlabeled UA (Standards or samples) with the pre-coated antibody specific to UA. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of UA in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of UA in the sample.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/general-elisa-kits/uric-acid-ua/189663>