

## Human 8-Oxoguanine DNA Glycosylase (hOGG1) ELISA Kit | hOGG1 elisa kit

Human 8-Oxoguanine DNA Glycosylase (hOGG1) ELISA Kit

CATALOG #  
**AAA201996**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

8-Oxoguanine DNA Glycosylase (hOGG1); Human 8-Oxoguanine DNA Glycosylase (hOGG1) ELISA Kit; hOGG1 elisa kit

### PRODUCT SPECIFICATIONS

|                 |            |
|-----------------|------------|
| Reactivity      | Human      |
| Sequence Length | 341        |
| Sensitivity     | 1.0 nmol/L |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for hOGG1 elisa kit

Intended Uses: This HOGG1 ELISA kit is intended Laboratory for Research use only. The Stop Solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of HOGG1 in the sample, this HOGG1 ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus HOGG1 concentration. The concentration of HOGG1 in the samples is then determined by comparing the O.D. of the samples to the standard curve.

### NCBI AND UNIPROT INFORMATION

|                                  |                              |
|----------------------------------|------------------------------|
| NCBI GI #                        | 2598924                      |
| NCBI GeneID                      | 4968                         |
| NCBI Official Full Name          | 8-oxoguanine DNA glycosylase |
| NCBI Official Synonym Full Names | 8-oxoguanine DNA glycosylase |
| NCBI Official Symbol             | OGG1                         |
| NCBI Official Synonym Symbols    | HMMH; MUTM; OGH1; HOGG1      |
| NCBI Protein Information         | N-glycosylase/DNA lyase      |

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.