

Human Excision Repair Cross-Complementation Group 1 (ERCC1) ELISA Kit | ERCC1 elisa kit

Human Excision Repair Cross-Complementation Group 1 (ERCC1) ELISA Kit

CATALOG # AAA203036	REACTIVITY Human
-------------------------------	----------------------------


PRODUCT SYNONYMS

Excision Repair Cross-Complementation Group 1 (ERCC1); Human Excision Repair Cross-Complementation Group 1 (ERCC1) ELISA Kit; ERCC1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Sequence Length	271
Samples	Serum, plasma, cell culture supernate and other biological fluids
Assay Type	Sandwich
Detection Range	0.5ng/ml-8ng/ml
Sensitivity	0.1 ng/ml

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ERCC1 elisa kit	Intended Use: The Stop Solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of ERCC1 in the sample, This ERCC1 ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus ERCC1 concentration. The concentration of ERCC1 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	1430349972
-----------	------------

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/excision-repair-cross-complementation-group-1-ercc1/203036>