

## Human Cross Linked C-Telopeptide Of Type III Collagen ELISA Kit

Human Cross Linked C-Telopeptide Of Type III Collagen ELISA Kit

CATALOG #  
**AAA203836**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

Cross Linked C-Telopeptide Of Type III Collagen; Human Cross Linked C-Telopeptide Of Type III Collagen ELISA Kit; Cross Linked C-Telopeptide Of Type III Collagen elisa kit

### PRODUCT SPECIFICATIONS

|                 |                                                                   |
|-----------------|-------------------------------------------------------------------|
| Reactivity      | Human                                                             |
| Samples         | Serum, plasma, cell culture supernate and other biological fluids |
| Assay Type      | Sandwich                                                          |
| Detection Range | 2ng/ml-32ng/ml                                                    |
| Sensitivity     | 0.1 ng/ml                                                         |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for Cross Linked C-Telopeptide Of Type III Collagen elisa kit

Intended Use: The Stop Solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of Cross Linked C-Telopeptide Of Type III Collagen in the sample, This Cross Linked C-Telopeptide Of Type III Collagen Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus Cross Linked C-Telopeptide Of Type III Collagen concentration. The concentration of Cross Linked C-Telopeptide Of Type III Collagen in the samples is then determined by comparing the O.D. of the samples to the standard curve.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/cross-linked-c-telopeptide-of-type-iii-collagen/203836>