

Human Transient Receptor Potential Cation Channel Subfamily V, Member 1 (TRPV1) ELISA Kit | TRPV1 elisa kit

Human Transient Receptor Potential Cation Channel Subfamily V, Member 1 (TRPV1) ELISA Kit

CATALOG # AAA20415	REACTIVITY Human
------------------------------	----------------------------


PRODUCT SYNONYMS

Transient Receptor Potential Cation Channel Subfamily V; Member 1 (TRPV1); Human Transient Receptor Potential Cation Channel Subfamily V; Member 1 (TRPV1) ELISA Kit; VR1; OTRPC1; Vanilloid Receptor Subtype 1; Capsaicin receptor; Osm-9-like TRP channel 1; Vanilloid receptor 1; TRPV1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of TRPV1. No significant cross-reactivity or interference between TRPV1 and analogues was observed.
Sequence Length	839
Assay Type	Quantitative Sandwich
Samples	Human tissue homogenates, cell lysates and other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.121ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level TRPV1 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level TRPV1 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TRPV1 elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to TRPV1. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to TRPV1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain TRPV1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of TRPV1 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of TRPV1 in human tissue homogenates, cell lysates and other biological fluids.
Product Categories/Family for TRPV1 elisa kit	Signal Transduction

NCBI AND UNIPROT INFORMATION

NCBI GI #	74315354
NCBI GeneID	7442
NCBI Accession #	NP_542437.2
NCBI GenBank Nucleotide #	NM_080706.3
UniProt Accession #	Q8NER1; Q3LU47; Q9H0G9; Q9H303; Q9H304; Q9NQ74; Q9NY22; A2RUA9
Molecular Weight	94,211 Da
NCBI Official Full Name	transient receptor potential cation channel subfamily V member 1
NCBI Official Synonym Full Names	transient receptor potential cation channel, subfamily V, member 1
NCBI Official Symbol	TRPV1
NCBI Official Synonym Symbols	VR1
NCBI Protein Information	transient receptor potential cation channel subfamily V member 1
UniProt Protein Name	Transient receptor potential cation channel subfamily V member 1
UniProt Gene Name	TRPV1
UniProt Synonym Gene Names	VR1; TrpV1; OTRPC1
UniProt Entry Name	TRPV1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.