

Human Absent In Melanoma 2 (AIM2) ELISA Kit | AIM2 elisa kit

Human Absent In Melanoma 2 (AIM2) ELISA Kit

CATALOG #
AAA20451

REACTIVITY
Human


PRODUCT SYNONYMS

Absent In Melanoma 2 (AIM2); Human Absent In Melanoma 2 (AIM2) ELISA Kit; PYHIN4; Interferon-inducible protein AIM2; AIM2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of AIM2. No significant cross-reactivity or interference between AIM2 and analogues was observed.
Sequence Length	343
Assay Type	Quantitative Sandwich
Samples	Human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.156-10ng/mL
Sensitivity	< 0.061ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level AIM2 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level AIM2 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AIM2 elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to AIM2. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to AIM2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain AIM2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of AIM2 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of AIM2 in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.
Product Categories/Family for AIM2 elisa kit	Tumor immunity

NCBI AND UNIPROT INFORMATION

NCBI GI #	119573180
NCBI GeneID	9447
UniProt Accession #	O14862; Q5T3V9; Q96FG9; A8K7M7
Molecular Weight	38,954 Da
NCBI Official Full Name	absent in melanoma 2
NCBI Official Synonym Full Names	absent in melanoma 2
NCBI Official Symbol	AIM2
NCBI Official Synonym Symbols	PYHIN4
NCBI Protein Information	interferon-inducible protein AIM2
UniProt Protein Name	Interferon-inducible protein AIM2
UniProt Gene Name	AIM2
UniProt Entry Name	AIM2_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/absent-in-melanoma-2-aim2/20451>