

Mouse GLUT1 ELISA Kit | GLUT1 elisa kit

Mouse GLUT1 (Glucose Transporter 1) ELISA Kit

CATALOG #
AAA21572

REACTIVITY
Mouse


PRODUCT SYNONYMS

GLUT1; Mouse GLUT1 (Glucose Transporter 1) ELISA Kit; GLUT1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	This kit recognizes Mouse GLUT1 in samples. No significant cross-reactivity or interference between Mouse GLUT1 and analogues was observed.
Sequence Length	130
Samples	Serum, plasma and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	0.31-20ng/mL
Sensitivity	0.19ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Mouse GLUT1 were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Mouse GLUT1 were tested on 3 different plates, 20 replicates in each plate.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for GLUT1 elisa kit	Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Mouse GLUT1 concentrations in serum, plasma and other biological fluids. Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Mouse GLUT1. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Mouse GLUT1 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Mouse GLUT1, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The OD value is proportional to the concentration of Mouse GLUT1. You can calculate the concentration of Mouse GLUT1 in the samples by comparing the OD of the samples to the standard curve.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	14268550
NCBI GeneID	6513
UniProt Accession #	P11166; O75535; Q147X2; A8K9S6; B2R620; D3DPX0
Molecular Weight	54,084 Da
NCBI Official Full Name	glucose transporter-like protein I, partial
NCBI Official Synonym Full Names	solute carrier family 2 (facilitated glucose transporter), member 1
NCBI Official Symbol	SLC2A1
NCBI Official Synonym Symbols	PED; DYT9; GLUT; DYT17; DYT18; EIG12; GLUT1; HTLVR; GLUT-1; GLUT1DS
NCBI Protein Information	solute carrier family 2, facilitated glucose transporter member 1; glucose transporter type 1, erythrocyte/brain; hepG2 glucose transporter; human T-cell leukemia virus (I and II) receptor; receptor for HTLV-1 and HTLV-2
UniProt Protein Name	Solute carrier family 2, facilitated glucose transporter member 1
UniProt Gene Name	SLC2A1
UniProt Synonym Gene Names	GLUT1; GLUT-1
UniProt Entry Name	GTR1_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.