

Human 2',5'-Oligoadenylate Synthetase 2 (OAS2) RTU ELISA Kit | OAS2 rtu elisa kit

Human 2',5'-Oligoadenylate Synthetase 2 (OAS2) ELISA Kit

CATALOG #
AAA217524

REACTIVITY
Human


PRODUCT SYNONYMS

2'; 5'-Oligoadenylate Synthetase 2 (OAS2); Human 2'; 5'-Oligoadenylate Synthetase 2 (OAS2) ELISA Kit; p69 OAS / p71 OAS; OAS2 rtu elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of OAS2. No significant cross-reactivity or interference between OAS2 and analogues was observed.
Sequence Length	172
Samples	Serum, plasma, tissue homogenates or other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	3.12-200ng/mL
Sensitivity	< 1.24ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level OAS2 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level OAS2 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kits: All the reagents should be kept at 4 degree C upon receipt. For opened kits: Once the kit is opened, the remaining reagents still need to be stored according to the above storage conditions. In addition, return the unused wells to the foil pouch containing the desiccant pack and reseal along entire edge of zip-seal.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OAS2 rtu elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for the in vitro quantitative measurement of OAS2 in human serum, plasma, tissue homogenates or other biological fluids. Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to OAS2. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to OAS2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain OAS2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of OAS2 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	74229021
NCBI GeneID	4939
NCBI Accession #	NP_001027903.1
NCBI GenBank Nucleotide #	NM_001032731.1
UniProt Accession #	P29728; Q6PJ33; Q86XX8; A8K9T1
Molecular Weight	19,517 Da
NCBI Official Full Name	2'-5'-oligoadenylate synthase 2 isoform 3
NCBI Official Synonym Full Names	2'-5'-oligoadenylate synthetase 2
NCBI Official Symbol	OAS2
NCBI Protein Information	2'-5'-oligoadenylate synthase 2
UniProt Protein Name	2'-5'-oligoadenylate synthase 2
UniProt Gene Name	OAS2
UniProt Synonym Gene Names	p69OAS / p71OAS

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-rtu-elisa-kits/25-oligoadenylate-synthetase-2-oas2/217524>