

Human IgA2 ELISA Kit | IgA2 elisa kit

Human IgA2 (Immunoglobulin A2) ELISA Kit

CATALOG #
AAA21900

REACTIVITY
Human


PRODUCT SYNONYMS

IgA2; Human IgA2 (Immunoglobulin A2) ELISA Kit; IgA2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes Human IgA2 in samples. No significant cross-reactivity or interference between Human IgA2 and analogues was observed.
Sequence Length	1887
Samples	Serum and plasma, please inquire if your samples are other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	1.56-100ng/mL
Sensitivity	0.94ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human IgA2 were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human IgA2 were tested on 3 different plates, 20 replicates in each plate.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IgA2 elisa kit	Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Human IgA2 concentrations in serum and plasma, please inquire if your samples are other biological fluids. Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human IgA2. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human IgA2 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human IgA2, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 +/- 2 nm. The OD value is proportional to the concentration of Human IgA2. You can calculate the concentration of Human IgA2 in the samples by comparing the OD of the samples to the standard curve.
--	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	89892359
Molecular Weight	210,554 Da
NCBI Official Full Name	Iga2
UniProt Protein Name	Iga2
UniProt Gene Name	iga2
UniProt Entry Name	Q19V48_HAEIF

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/iga2/21900>