

Human Acetyl-CoA Carboxylase ELISA Kit | ACC elisa kit

Human ACC (Acetyl-CoA Carboxylase) ELISA Kit

CATALOG #
AAA22031

REACTIVITY
Human


PRODUCT SYNONYMS

Acetyl-CoA Carboxylase; Human ACC (Acetyl-CoA Carboxylase) ELISA Kit; ACC elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This kit recognizes Human ACC in samples. No significant cross-reactivity or interference between Human ACC and analogues was observed
Sequence Length	2346
Assay Type	Quantitative Sandwich
Samples	Serum, plasma and other biological fluids
Detection Range	1.56-100 ng/mL
Sensitivity	0.94 ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Human ACC were tested 20 times on one plate, respectively.
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Human ACC were tested on 3 different plates, 20 replicates in each plate, respectively.
Preparation and Storage	Store at 4 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ACC elisa kit	Principle of the Assay: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human ACC. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human ACC and Avidin- Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human ACC, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human ACC. You can calculate the concentration of Human ACC in the samples by comparing the OD of the samples to the standard curve. Intended Uses: This ELISA kit applies to the in vitro quantitative determination of Human C3b concentrations in serum, plasma and other biological fluids.
---	--

NCBI AND UNIPROT INFORMATION

NCBI GI #	6006405
NCBI GeneID	281590
UniProt Accession #	Q9TTS3
Molecular Weight	265,303 Da
NCBI Official Full Name	acetyl-CoA-carboxylase
NCBI Official Symbol	ACACA
NCBI Official Synonym Symbols	ACC1; ACCA
NCBI Protein Information	acetyl-CoA carboxylase 1
UniProt Protein Name	Acetyl-CoA carboxylase 1
UniProt Gene Name	ACACA
UniProt Synonym Gene Names	ACAC; ACCA; ACC1
UniProt Entry Name	ACACA_BOVIN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/acc-human-elisa-kits/acetyl-coacarboxylase/22031>