

Human Placental Protein 13 (PP13) ELISA Kit | PP13 elisa kit

Human Placental Protein 13 (PP13) ELISA Kit

CATALOG #
AAA23004

REACTIVITY
Human


PRODUCT SYNONYMS

Placental Protein 13 (PP13); Human Placental Protein 13 (PP13) ELISA Kit; LGALS13; PLAC8; PP13; GAL13; Galectin 13; Galactoside-Binding Soluble Lectin 13; Placental tissue protein 13; PP13 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of PP13. No significant cross-reactivity or interference between PP13 and analogues was observed.
Sequence Length	139
Assay Type	Quantitative Sandwich
Samples	Human serum, plasma, tissue homogenates and other biological fluids
Detection Range	15.6-1,000pg/mL
Sensitivity	< 5.5pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level PP13 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level PP13 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PP13 elisa kit	Principle of the Assay: The microplate provided in this kit has been pre-coated with an antibody specific to PP13. Standards or samples are then added to the appropriate microplate wells with a biotin-conjugated antibody specific to PP13. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain PP13, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of PP13 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of PP13 in human serum, plasma, tissue homogenates and other biological fluids.
--	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	6650760
NCBI GeneID	29124
UniProt Accession #	Q9UHV8; C5HZ15
Molecular Weight	16,119 Da
NCBI Official Full Name	placental protein 13
NCBI Official Synonym Full Names	galectin 13
NCBI Official Symbol	LGALS13
NCBI Official Synonym Symbols	PP13; GAL13; PLAC8
NCBI Protein Information	galactoside-binding soluble lectin 13
UniProt Protein Name	Galactoside-binding soluble lectin 13
UniProt Gene Name	LGALS13
UniProt Synonym Gene Names	PLAC8; Gal-13; PP13; Placental protein 13

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.