

Sheep 1, 3-beta-D-glucan ELISA Kit | BDG elisa kit

Sheep 1, 3-beta-D-glucan ELISA Kit

CATALOG #
AAA246440

REACTIVITY
Sheep


PRODUCT SYNONYMS

1; 3-beta-D-glucan; Sheep 1; 3-beta-D-glucan ELISA Kit; 3-b-D-glucan ELISA Kit; BDG elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Sheep
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Sensitivity	1.0 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for BDG elisa kit

Intended Uses: This betaDG ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Sheep betaDG. This ELISA kit for research use only!

Principle of the Assay||betaDG ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-betaDG antibody and an betaDG-HRP conjugate. The assay sample and buffer are incubated together with betaDG-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the betaDG concentration since betaDG from samples and betaDG-HRP conjugate compete for the anti-betaDG antibody binding site. Since the number of sites is limited, as more sites are occupied by betaDG from the sample, fewer sites are left to bind betaDG-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The betaDG concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.