

Porcine Crosslaps, Cr ELISA Kit | Cr elisa kit

Porcine Crosslaps, Cr ELISA Kit

CATALOG #
AAA246574

REACTIVITY
Porcine


PRODUCT SYNONYMS

Crosslaps; Cr; Porcine Crosslaps; Cr ELISA Kit; Cr elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Specificity	This assay has high sensitivity and excellent specificity for detection of CR. No significant cross-reactivity or interference between CR and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between CR and all the analogues, therefore, cross reaction may still exist in some cases.
Assay Type	Quantitative Competitive
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Sensitivity	1.0 pg/mL
Intended Uses	This Crosslaps ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine Crosslaps. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Cr elisa kit	Principle of the Assay: CR ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-CR antibody and an CR-HRP conjugate. The assay sample and buffer are incubated together with CR-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the CR concentration since CR from samples and CR-HRP conjugate compete for the anti-CR antibody binding site. Since the number of sites is limited, as more sites are occupied by CR from the sample, fewer sites are left to bind CR-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The CR concentration in each sample is interpolated from this standard curve. Intended Uses: This CR ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine CR. This ELISA kit for research use only, not for therapeutic or test applications!
--	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/porcine-elisa-kits/crosslaps-cr/246574>