

Canine acylated ghrelin (AG) ELISA Kit | AG elisa kit

Canine acylated ghrelin (AG) ELISA Kit

CATALOG #
AAA246623

REACTIVITY
Canine


PRODUCT SYNONYMS

acylated ghrelin (AG); Canine acylated ghrelin (AG) ELISA Kit; AG elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AG elisa kit

Intended Uses: This AG ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine AG.

Principle of the Assay: AG ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-AG antibody and an AG-HRP conjugate. The assay sample and buffer are incubated together with AG-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the AG concentration since AG from samples and AG-HRP conjugate compete for the anti-AG antibody binding site. Since the number of sites is limited, as more sites are occupied by AG from the sample, fewer sites are left to bind AG-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The AG concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.