

Bovine calpain 1 (CAPN1) ELISA Kit | CAPN1 elisa kit

Bovine calpain 1 (CAPN1) ELISA Kit

CATALOG #
AAA246779

REACTIVITY
Bovine


PRODUCT SYNONYMS

calpain 1 (CAPN1); Bovine calpain 1 (CAPN1) ELISA Kit; CAPN1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



ADDITIONAL INFORMATION

Related Product Information for CAPN1 elisa kit

Intended Uses: This CAPN1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Bovine CAPN1. This ELISA kit for research use only!

Principle of the Assay: CAPN1 ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-CAPN1 antibody and an CAPN1-HRP conjugate. The assay sample and buffer are incubated together with CAPN1-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the CAPN1 concentration since CAPN1 from samples and CAPN1-HRP conjugate compete for the anti-CAPN1 antibody binding site. Since the number of sites is limited, as more sites are occupied by CAPN1 from the sample, fewer sites are left to bind CAPN1-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The CAPN1 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

