

Bovine Ubiquitin Carboxy-Terminal Hydrolases L1 ELISA Kit | UCH-L1 elisa kit

Bovine Ubiquitin Carboxy-Terminal Hydrolases L1 ELISA Kit

CATALOG #
AAA247536

REACTIVITY
Bovine


PRODUCT SYNONYMS

Ubiquitin Carboxy-Terminal Hydrolases L1; Bovine Ubiquitin Carboxy-Terminal Hydrolases L1 ELISA Kit; UCH-L1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Specificity	This assay has high sensitivity and excellent specificity for detection of USPL1. No significant cross-reactivity or interference between USPL1 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between USPL1 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 ng/mL
Preparation and Storage	Store at 2-8 degree C. The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for UCH-L1 elisa kit

Intended Uses: This USPL1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Bovine USPL1. This ELISA kit for research use only, not for therapeutic or test applications!

Principle of the Assay: USPL1 ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-USPL1 antibody and an USPL1-HRP conjugate. The assay sample and buffer are incubated together with USPL1-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the USPL1 concentration since USPL1 from samples and USPL1-HRP conjugate compete for the anti-USPL1 antibody binding site. Since the number of sites is limited, as more sites are occupied by USPL1 from the sample, fewer sites are left to bind USPL1-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The USPL1 concentration in each sample is interpolated from this standard curve.

Product Categories/Family for UCH-L1 elisa kit

Immunology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/uch-l1-bovine-ovine-elisa-kits/ubiquitin-carboxy-terminal-hydrolases-l1/247536>