

Porcine anti-Interferon omega IgG antibody ELISA Kit | IFNomega IgG-AB elisa kit

Porcine anti-Interferon omega IgG antibody ELISA Kit

CATALOG #
AAA247713

REACTIVITY
Porcine


PRODUCT SYNONYMS

anti-Interferon omega IgG antibody; Porcine anti-Interferon omega IgG antibody ELISA Kit; IFNomega IgG-AB elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Specificity	This assay has high sensitivity and excellent specificity for detection of IFN-omega IgG-Ab. No significant cross-reactivity or interference between IFN- omega IgG-Ab and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between IFN-omega IgG-Ab and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IFNomega IgG-AB elisa kit

Intended Uses: This IFN-omega IgG-Ab ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine IFN-omega IgG-Ab. This ELISA kit for research use only, not for therapeutic or test applications!

Principle of the Assay: IFN-omega IgG-Ab ELISA kit applies the competitive enzyme immunoassay technique utilizing a IFN-omega IgG antigen and an IFN-omega IgG-Ab-HRP conjugate. The assay sample and buffer are incubated together with IFN-omega IgG-Ab-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the IFN-omega IgG-Ab concentration since IFN-omega IgG-Ab from samples and IFN-omega IgG-Ab-HRP conjugate compete for the IFN-omega IgG antigen binding site. Since the number of sites is limited, as more sites are occupied by IFN-omega IgG-Ab from the sample, fewer sites are left to bind IFN-omega IgG-Ab-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The IFN-omega IgG-Ab concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

