

Fish Cyclooxygenase 2 ELISA Kit | COX2 elisa kit

Fish Cyclooxygenase 2 ELISA Kit

CATALOG #
AAA247760

REACTIVITY
Fish


PRODUCT SYNONYMS

Cyclooxygenase 2; Fish Cyclooxygenase 2 ELISA Kit; COX2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Fish
Specificity	This assay has high sensitivity and excellent specificity for detection of COX2. No significant cross-reactivity or interference between COX2 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between COX2 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for COX2 elisa kit	Intended Uses: This COX2 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Fish COX2. This ELISA kit for research use only, not for therapeutic or test applications! Principle of the Assay: COX2 ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-COX2 antibody and an COX2-HRP conjugate. The assay sample and buffer are incubated together with COX2-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the COX2 concentration since COX2 from samples and COX2-HRP conjugate compete for the anti-COX2 antibody binding site. Since the number of sites is limited, as more sites are occupied by COX2 from the sample, fewer sites are left to bind COX2-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The COX2 concentration in each sample is interpolated from this standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/cox2-fish-elisa-kits/cyclooxygenase-2/247760>